



RADIATION PROTECTION SURVEY

Facility Info	<i>Name:</i>	Flight Dental Systems	
	<i>Address:</i>	21 Kenview Blvd Unit 9 Brampton, ON	
	<i>Country:</i>	Canada	<i>Telephone:</i> 905-799-0517

Equipment	<i>Room No:</i>	NA	<i>Type:</i>	Portable Handheld X-Ray
	<i>Manuf:</i>	Flight Dental Systems	<i>Model:</i>	X-Vision
	<i>System SN:</i>	U00309/02929	<i>DOM:</i>	Not found

Cone	<i>Model:</i>	XR-3800BL	<i>SN:</i>	U00309BL
-------------	---------------	-----------	------------	----------

Housing	<i>Model:</i>	XR-3800	<i>SN:</i>	U00309/02929
----------------	---------------	---------	------------	--------------

Tube Insert	<i>Model:</i>	Canon D-045	<i>SN:</i>	Not found
--------------------	---------------	-------------	------------	-----------

Survey Information	<i>Date:</i>	February 28, 2022		
	<i>Detector:</i>	RaySafe	<i>Model:</i>	452
	<i>SN:</i>	278094	<i>Last Cal Date:</i>	6/2/2021

METHOD:


A radiation detector was used to measure stray radiation in areas immediately surrounding the x-ray unit. A 16 cm CTDI PMMA Phantom was used as a scatter medium. The technique used was 60 kV, 2 mA, 2.00 seconds.

Christopher A Lease, MS, DABR
Board Certified Medical Physicist

Technique: 60 kV @ 4 mAs (2 mA, 2.0 seconds)
 Phantom: 16cm CTDI PMMA
 Background: 12 μ R/hr
 Scatter at 0.3 m with no phantom: 2.2 mR/hr

Backscatter shield primary beam attenuation evaluation

Technique: 80 kVp @ 20 mAs, 100 cm Source-to-Detector Dist
 Measured Exposure (without shield): 123.6 mR
 Measured Exposure (with shield): 3.859 mR
 Result: 97% of primary beam attenuated by backscatter shield



*30 cm is the minimum object scatter-to-operator distance measurable